

Compliance Declaration

We, Shenzhen Dongdixin Technology Co., Ltd., address in No.3. Building Xilibaimang Xusheng Industrial Estate Nanshan, Shenzhen P. R. China, 518108 declare that Thermometer TH2030, TH2071, TH1009, TH1091 comply with the following standard:

- IEC 60601-1:2012 Medical electrical equipment Part 1. General requirements for basic safety and essential performance
- IEC60601-1-11:2015 General requirements for basic safety and essential performance-Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
- IEC 60601-1-2:2007 Medical electrical equipment -- Part 1-2: General requirements for basic safety and essential performance -- Collateral standard: Electromagnetic compatibility - Requirements and tests
- ISO 80601-2-56:2009 Medical electrical equipment --Part 2-56:Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement

The declaration is valid from Mar. 6th 2018 to Mar. 5th 2019.

Shenzhen 2018.3.7

Place, date

Shenzhen Dongdixin Technology Co.,Ltd

Title: VP

Name: Jianping Kang